



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

January 15, 2014

SIRONA DENTAL INCORPORATED
STEVE SUTTON
4835 SIRONA DR SUITE 100
CHARLOTTE NC 28273

SIRONA DENTAL SYSTEMS, GMBH
STEVE SUTTON
FABRIKSTRASSE 31
BENSHEIM D-64625
GERMANY

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SIRONA DENTAL SYSTEMS, GMBH
Product Name: SIRONA C8+
Model Number(s): D3417
Product File No: 20130383

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of January 2019.

This approval supersedes the approval issued on February 17, 2009 under product file number 20090083.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail